

1. Einleitung

Normales Sehen ist ein entscheidender Faktor für die Interaktion eines Organismus mit seiner Umwelt. Für die meisten Tiere stellt der Verlust der Sehkraft eine wesentliche Behinderung dar. Um sich vor Raubtieren zu schützen, hat sich das Sehvermögen von Pferden an die Erkennung von Bewegungen bei schlechten Lichtbedingungen angepasst [1]. Zudem ermöglichen die laterale Position der Augen am Pferdekopf und die horizontal elliptische Pupille dem Pferd ein großes horizontales Sichtfeld von circa 350 Grad [2]. Ein gutes Sehvermögen ist jedoch nicht nur für Wildpferde, sondern auch für die domestizierten Pferde, welche als Freizeit-, Turnier-, Arbeits- oder Zuchtpferde genutzt werden, von größter Bedeutung [1]. In Ihrer Sehkraft eingeschränkte Pferde sind wertgemindert, zum Teil gefährlich und prädisponiert für Selbstverletzungen. Beidseitig erblindete Pferde werden häufig euthanasiert [3]. Beim Pferd gilt als Hauptursache für eine Erblindung die Katarakt [4]. Als Katarakt oder grauer Star wird jede Trübung der Augenlinse bezeichnet [5]. Die Transparenz der Linse wird durch verschiedene Faktoren gewährleistet. Dazu gehören: die geringe Zytoplasmadichte aufgrund des Fehlens von intrazellulären Organellen einschließlich des Zellkerns in den Linsenfasern, der geringe zytoplasmatische Brechungsindex sowie die streng ausgerichtete Gitteranordnung der Linsenfasern. Kommt es zu Veränderungen im Linsenmetabolismus, wird die Linse trüb und eine Katarakt entsteht. Solche Veränderungen werden durch Störungen im Proteingehalt der Linse, an den metabolischen Pumpen, in der Ionenkonzentration und an der antioxidativen Aktivität verursacht [6].

Katarakte können nach verschiedenen Kriterien eingeteilt werden: anatomische Lage in der Linse, physikalische Erscheinung, mutmaßliche Ätiologie oder Entwicklungsstadium. Außerdem werden sie unterteilt in angeborene und erworbene Katarakte [1]. Beim Pferd sind 0,5-5,3% aller angeborenen Defekte Augendefekte und davon sind 33,6%-35,5% Katarakte [3, 7]. Eine Vererbung konnte bei den Rassen Vollblut, Quarter Horse und Morgan Horse nachgewiesen werden [4]. Katarakte gelten als die häufigste Erkrankung der Linse und sollen bei 5%-7% aller Pferde mit "gesunden" Augen vorkommen. Als häufigste Ursache für erworbene bzw. sekundäre Katarakte bei Pferden gilt die Uveitis, z. B. in Form der equine recurrent uveitis (ERU) oder auch Mondblindheit genannt [3, 8]. Zu den Katarakt auslösenden Faktoren



gehören weiterhin systemische oder metabolische Erkrankungen, UV- oder Mikrowellenstrahlung bzw. einige Arzneimittel und Umweltgifte [1]. *Therapie*

Derzeit gibt es keine Medikamente, welche eine Therapie oder Prophylaxe von Katarakten ermöglichen [3, 9]. Die Katarakt-Operation ist die einzige Therapie, welche die Sehfähigkeit wieder herstellen kann [10]. Seit dem 18. Jahrhundert wird die chirurgische Behandlung von Katarakten beschrieben. Damals wurde die erkrankte Linse mithilfe eines einer Lanzette ähnlichen Instruments in den Glaskörper geschoben. Beim Pferd wurden weitere Operationstechniken zur Entfernung der Linse entwickelt. Dazu gehören die Aspiration der weichen kataraktösen Linse von Fohlen, die intrakapsuläre Extraktion, die extrakapsuläre Extraktion und die Phacoemulsifikation mit Aspiration. Letztgenannte setzte sich gegenüber den anderen Techniken durch und ist heute eine bevorzugte und weit verbreitete Operationstechnik [1, 11]. Häufig werden die Pferde aphak (ohne Linse) gelassen, woraus eine starke Weitsichtigkeit resultiert. Aphake Pferde haben im Durchschnitt eine Fehlsichtigkeit von +9,5 D [3]. Verschiedene Studien zeigen, dass diese Pferde trotz ihrer Fehlsichtigkeit normales Verhalten zeigen und in den verschiedenen Nutzungsrichtungen (z. B. Westernpferd, Freizeitpferd, Turnierpferd) verwendet werden können. Jedoch werden aber auch visuelle Beeinträchtigungen wie ein reduziertes Nachtsehen, eine verminderte Kontrastempfindlichkeit und eine erhöhte Blendeempfindlichkeit festgestellt [12-15]. Um die Fehlsichtigkeit zu korrigieren, gibt es für Pferde Intraokularlinsen (IOL) in zwei verschiedenen Stärken (14 D und 18 D). Beim Pferd findet die Implantation von IOLs anders als bei Menschen, Hunden und Katzen noch nicht routinemäßig statt und wird eher selten durchgeführt. In der humanen Augenheilkunde gibt es eine große Auswahl an IOLs in verschiedenen Stärken und für jeden Patienten wird vor der Kataraktoperation die individuell passende Linse ermittelt [16]. Beim Hund und bei der Katze wurden biometrische und keratometrische Messungen vorgenommen und mit Hilfe theoretischer Formeln konnten für diese Tierarten kommerziell erhältliche Standard-IOL entwickelt werden [17, 18].

In der Humanmedizin wurden zur Berechnung der IOL-Stärke verschiedene Formeln entwickelt. Diese können in theoretische Formeln und empirische Regressionsformeln unterteilt werden. Um mit Hilfe dieser Formeln den Brechwert einer IOL bestimmen zu können, benö-

Einleitung

tigt man Werte, welche durch keratometrische und biometrische Messungen ermittelt werden. Die Kenntnis über die optischen Maße des Pferdeauges ist die Voraussetzung für die Bestimmung der IOL-Stärke beim Pferd. Die Messungen dafür müssen so genau wie möglich durchgeführt werden. In früheren Studien [16, 19-21] wurden bei Pferden die verschiedenen Werte (Hornhautradius bzw. Brechkraft der Hornhaut, Vorderkammertiefe, Linsendicke, Axiallänge) ermittelt, jedoch wurden gerade bei der Ultraschallmessung Geräte verwendet, die auf das humane Auge eingestellt waren. Für biometrische Messungen ist der A-Mode Ultraschall die Methode der Wahl [22]. Die Länge der intraokularen Strukturen ergibt sich aus der Formel l=ct, wobei l die Länge, c die Ultraschallgeschwindigkeit und t die Laufzeit des Ultraschalls sind. Zu beachten ist jedoch, dass in den verschiedenen Geweben unterschiedliche Ultraschallgeschwindigkeiten vorherrschen. Beim Menschen betragen diese für Kammerwasser und Glaskörper 1532 m/s und für gesundes Linsengewebe 1641 m/s [23]. In früheren Studien [24-28] konnte festgestellt werden, dass die Ultraschallgeschwindigkeit im Linsengewebe bei verschiedenen Tierarten und dem Menschen unterschiedlich ist.

Ziel dieser Studie ist es, die Ultraschallgeschwindigkeit in equinem Kammerwasser, Glaskörper und Linse zu bestimmen und den Einfluss von Alter, Temperatur, Geschlecht und Zeit nach der Enukleation zu untersuchen. Entsprechende Umrechnungsfaktoren sollen berechnet werden. In einem zweiten Teil sollen an gesunden Pferdeaugen die Refraktion mittels Retinoskopie, die vordere Augenkammertiefe, die Linsendicke, die Glaskörperlänge und die Axiallänge des Auges mit A-Mode Ultraschalluntersuchung unter Verwendung der ermittelten Ultraschallgeschwindigkeiten für die verschiedenen Gewebe im Pferdeauge und die Brechkraft der Hornhaut mit einem Videokeratometer bestimmt werden. Außerdem soll der Einfluss von Rasse, Widerristhöhe, Geschlecht und Alter ermittelt und die Abhängigkeit der Werte zwischen dem linken und rechten Auge untersucht werden. Die Werte sollen genutzt werden, um mit Hilfe der Formeln nach Binkhorst und Retzlaff die Stärke von Intraokularlinsen für Pferde zu bestimmen und diese mit Ergebnissen von IOL- Implantationen aus früheren Studien zu vergleichen.

2. Ergebnisse

2.1 Publikation I:

Evaluation of ultrasound velocity in enucleated equine aqueous humor, lens and vitreous body

Published in: BMC Veterinary Research, 2014; 10:250

Received: 20 June 2014; Accepted: 6 October 2014; published online: 14 October 2014

U. Meister, B. Ohnesorge, D. Körner, M.H. Boevé Stiftung Tierärztliche Hochschule Hannover, Klinik für Pferde Bünteweg 9, 30559 Hannover, Germany

Abstract:

Background: Sonographic ophthalmic examinations have become increasingly important in veterinary medicine. If the velocity of ultrasound in ocular tissues is known, the A-mode ultrasound method may be used to determine the axial intraocular distances, such as anterior chamber depth, lens thickness, axial length of the vitreous and axial globe length, which are required for intraocular lens (IOL) power calculations. To the authors' knowledge, the velocity of ultrasound in the ocular tissues of the horse was not previously determined. In the present study, 33 lenses, 29 samples of aqueous and 31 of vitreous from 35 healthy equine eyes have been examined. The corresponding ultrasound velocities are reported in dependence of age, temperature, gender and elapsed time after enucleation.

Results: The velocity of ultrasound at 36°C in equine aqueous, lens and vitreous are 1529±10 m/s, 1654±29 m/s and 1527±16 m/s respectively, and the corresponding conversion factors are 0.998±0.007, 1.008±0.018 and 0.997±0.010. A linear increase of the speed of ultrasound with increasing temperature has been determined for aqueous and vitreous. No temperature dependence was found for the speed of ultrasound in the lens. The ultrasound veloci-



ty did not significantly differ (95 %) on the basis of gender, age or time after enucleation during the first 72 hours after death.

Conclusions: Compared to human eyes, the ultrasound velocity in equine lental tissue deviates by one percent. Therefore, axial length measurements obtained with ultrasound velocities for the human eye must be corrected using conversion factors. For the aqueous and vitreous, deviations are below one percent and can be neglected in clinical settings.

Keywords: Equine eye, Ultrasound velocity, Intraocular lens, Aqueous humor, Vitreous body

This article is published under license to BioMed Central Ltd. This is an Open Access article distributed under the terms of the Creative Commons Attribution License (<u>http://creativecommons.org/licenses/by/4.0</u>)

RESEARCH ARTICLE



Open Access

Evaluation of ultrasound velocity in enucleated equine aqueous humor, lens and vitreous body

Ulrike Meister^{1*}, Bernhard Ohnesorge¹, Daniel Körner² and Michael H Boevé¹

Abstract

Background: Sonographic ophthalmic examinations have become increasingly important in veterinary medicine. If the velocity of ultrasound in ocular tissues is known, the A-mode ultrasound method may be used to determine the axial intraocular distances, such as anterior chamber depth, lens thickness, axial length of the vitreous and axial globe length, which are required for intraocular lens (IOL) power calculations. To the authors' knowledge, the velocity of altrasound in the ocular tissues of the horse was not previously determined. In the present study, 33 lenses, 29 samples of aqueous and 31 of vitreous from 35 healthy equine eyes have been examined. The corresponding ultrasound velocities are reported in dependence of age, temperature, gender and elapsed time after enucleation.

Results: The velocity of ultrasound at 36°C in equine aqueous, lens and vitreous are 1529 ±10 m/s, 1654± 29 m/s and 1527 ±16 m/s respectively, and the corresponding conversion factors are 0.998±0.007, 1.008±0.018 and 0.997±0.010. A linear increase of the speed of ultrasound with increasing temperature has been determined for aqueous and vitreous. No temperature dependence was found for the speed of ultrasound in the lens. The ultrasound velocity did not significantly differ (95%) on the basis of gender, age or time after enucleation during the first 72 hours after death.

Conclusions: Compared to human eyes, the ultrasound velocity in equine lental tissue deviates by one percent. Therefore, axial length measurements obtained with ultrasound velocities for the human eye must be corrected using conversion factors. For the aqueous and vitreous, deviations are below one percent and can be neglected in clinical settings.

Keywords: Equine eye, Ultrasound velocity, Intraocular lens, Aqueous humor, Vitreous body

Background

In addition to a full ophthalmological examination, evaluation of the globe can be achieved by ultrasonography, which is used routinely in human medicine and has become increasingly important in veterinary medicine. Ultrasonographic examination is particularly important in the case of opacifications of the ocular media, e.g. by cataract. In veterinary ophthalmology, B-mode (brightness modulation) ultrasonography is used regularly. It provides a two-dimensional image that resembles a cross-section of the examined tissue. A-mode (amplitude modulation) ultrasound, e.g. for biometric measurements, is less commonly used. In human ophthalmology, the A-mode ultrasound is used before cataract surgery to determine the required dioptic power of the artificial

* Correspondence: Ulrike.Meister@t-online.de

Full list of author information is available at the end of the article



intraocular lens (IOL) and to monitor the postoperative state of the eye [1].

Accurate knowledge of the axial intraocular distances such as anterior chamber depth, lens thickness, axial length of the vitreous and axial globe length is necessary for IOL power calculations and the transversal size of the lens is important for the choice of the proper IOL size [2-4]. The method of choice is the biometric A-scan [5] which has previously been described for dogs [6], cats [7] and horses [8]. The intraocular distances (l) are calculated by determination of the travel time (t) of an ultrasonic wave with velocity (c) according to the formula l=ct. The majority of biometric studies of animal eyes are performed with an ultrasound machine that has been calibrated for humans. The travel time measurement is converted internally to a length measurement lusing the known ultrasound velocity c_h for the human eye. However, the ultrasound velocity differs among species and the true length *L* is given by $L = f \cdot l$ and $f = c_s/c_{lp}$ the

© 2014 Meister et al; [icensee BioMed Central Ltd, This is an Open Access article distributed under the terms of the Creative Commons Attribution License (http://creativecommons.org/licenses/bi/40), which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly credited. The Creative Commons Public Domain Dedication waiver (http://creativecommons.org/publicdomain/zero/1.0/) applies to the data made available in this article, unless otherwise stated.

12

¹Stiftung Tierärztliche Hochschule Hannover, Klinik für Pferde, Bünteweg 9, 30559 Hannover, Germany

conversion factor, where c_s is the ultrasound velocity of the examined species. It is therefore necessary to know the exact velocity of ultrasound propagation in the different ocular media for the species examined. In general, the ultrasound velocity depends upon different factors, including the ultrasound frequency and the type and temperature of the tissue examined.

The ultrasound velocity through ocular tissues has already been examined for humans and several animal species [2,9-15]. These studies indicate that the values for the velocity of ultrasound in lental tissue differ considerably among species, whereas no difference for the aqueous and vitreous is reported. The axial lens thickness is an important factor for the calculation of dioptic IOL power following the usual formulas of Retzlaff and Binkhorst and its accurate determination is, hence, very important [16]. A deviation of 50 m/s for the speed of ultrasound in a human lens already leads to uncertainty potential aberration of 0.5 D for the IOL power [17]. Therefore, considering different species, it is generally insufficient to estimate the IOL power by simply using the ultrasound velocity of the human lental tissue. The exact velocity for the horse should be used to achieve emmetropia in the eye. Visual soundness is very important for the pleasure and performance horse. A visually compromised horse may show abnormal behaviour and the rider faces an unpredictable, frightened animal [18].

To the authors' knowledge, the velocity of ultrasound in the ocular tissue of the horse has not been examined previously. The purpose of this study was to evaluate the velocity of ultrasound in equine aqueous humor, lens and vitreous in relationship to the factors age, temperature, gender and elapsed time after enucleation. The corresponding conversion factors have been calculated and implications for clinical ophthalmology are discussed.

Methods

Equine eyes were obtained from several slaughterhouses. Post mortem, every globe was enucleated and macroscopically examined. Afterwards the globes were immediately placed in a saline solution (0.9% Sodium chloride). All samples were cooled during transport. Only eyes without any signs of previous ophthalmic disease such as cataract or liquefaction of the vitreous body were examined. Depending on storage and the elapsed time after enucleation, the eyes were divided in three groups. Group A consists of 12 lenses, 10 samples of aqueous and 10 of vitreous, and the ocular tissues were stored at 8°C and examined within 12 hours after death. The eyes in group B were stored at 4°C and examined within 48 hours after death. This group consists of 11 lenses, 13 samples of aqueous and 12 of vitreous. Group C consists of eyes that were frozen at -18°C and examined within 72 hours after death, which amounts to 10 lenses, 6 samples of aqueous and 9 of vitreous. For some eyes, it was not possible to harvest enough aqueous or vitreous tissue for measurements. In addition, two lenses were damaged during the dissection process and excluded from the study.

Prior to examination, all eyes were adjusted to room temperature. In total, 35 healthy eyes were used for this study, out of which 17 eyes were from female and 18 from male horses. The age of the horses ranges from 6 to 22 years.

Preparation of the enucleated eyes started for group A within 12 hours, for group B within 48 hours and for group C within 72 hours post mortem by obtaining samples of the aqueous and vitreous, which were then stored in the sample-holders. The sclera was removed using scissors and the cornea was carefully separated to preserve the lens. Any remaining zonular material and vitreous was removed and the lens was placed in its respective sample holder. The examination of all ocular tissues occurred using three specially crafted sample holders, which consist of a copper frame to enable quick thermalisation and two Plexiglas plates, one of which was removable. The tissue sample was placed between the Plexiglas plates. The sample holder for the lens, aqueous and vitreous had sizes of 10 mm, 7 mm and 5 mm, respectively, in order to enable the tissue to be in full contact with the Plexiglas plates, thus securing a connection of the ultrasound signal between sample and holder without trapped air. By using sample holders with a fixed extent, the thickness of the tissue to be examined was determined accurately. In particular, measurement errors of the ultrasound velocity in the lens due to deformation for the length measurement, e.g. by the use of a manual calliper, could thus be avoided. The travel time of the ultrasonic wave was measured using the impulseecho method, which requires only a single transducer for ultrasound emission and registration. The ultrasound probe was directly connected to the first Plexiglas plate of the sample holder and a part of the ultrasound signal was transmitted through the sample (A-scan). Due to the difference of the refractive index between the Plexiglas and the sample, a clear signal could be assigned to the reflected wave corresponding to the transition between the first Plexiglas plate and the sample, as well as for the transition from the sample to the second Plexiglas plate. Therefore, a direct determination of the travel time (t) of the ultrasonic wave through the sample of thickness (d) was possible. Different setups for the measurement of the velocity of ultrasound such as the substitution technique or the double transmission technique have been described previously [9,10,19,20]. In the present study the double transmission technique was used and the velocity (c) is then provided by the formula c=2d/t.

The speed of ultrasound depends on the frequency of the ultrasound waves. For the present study, ultrasound

Page 2 of 6

equipment with a frequency range of 1 to 4 MHz was available. A 4 MHz transducer was used throughout. In order to obtain reference values, the ultrasound velocity for distilled water was measured for each sample holder separately and used to calibrate the experimental setup.

Temperature-dependent changes in the velocity of ultrasound were determined from $32^{\circ}C$ to $40^{\circ}C$ in $2^{\circ}C$ increments by placing the sample holders in a temperaturecontrolled water bath. For each measurement, five minutes of temperature adjustment were given.

For the velocities and conversion factors, the average and the standard deviation from the respective data subsets have been calculated. All hypotheses are evaluated using an appropriate variant of Student's *t*-test using a significance level of α =0.05. Regression curves are computed using the method of least squares and for linear regressions we give the value of Pearson's correlation coefficient *r*, which is related to the coefficient of determination R^2 by $R^2 = r^2$.

Results

Figures 1, 2 and 3 show the ultrasound velocity in the equine aqueous, lens and vitreous depending on the temperature for three separate sample groups that differ in time after enucleation and storage. The mean velocity for each tissue did not significantly depend (95%) on the time after enucleation and storage.

Hence, values for all sample groups have been averaged and the values for the velocity of ultrasound through aqueous, lens and vitreous a 36° C were 1529 ± 10 m/s, 1654 ± 29 m/s and 1527 ± 16 m/s and the conversion factors were 0.998 ± 0.007 , 1.008 ± 0.018 and 0.997 ± 0.010 .

A linear regression curve was plotted for the data averaged over all three sample groups. The result of this is



shown in Figure 4 and the value of the linear coefficient for lental tissue, which characterises the temperature dependence, is not significantly different from zero (95%). In contrast, the velocity of ultrasound in aqueous and vitreous increases significantly (95%) with increasing temperature and this was confirmed by a linear correlation. The corresponding regression formulas are depicted in Figure 4 and amounted to $c(m/s) = (1.23 \pm 0.13) \cdot T(^{\circ}C) + (1485 \pm 5)$ for the aqueous with correlation coefficient r=0.98 and $c(m/s) = (2.63 \pm 0.33) \cdot T(^{\circ}C) + (1430 \pm 12)$ for the vitreous with correlation coefficient r=0.98. No significant difference (95%) in the velocity of ultrasound between male and female individuals could be demonstrated. The







14



temperature-averaged ultrasound velocity through the lens for horses of different ages is shown in Figure 5. For each data point, the deviation from the mean velocity is not significant (95%) and this result is consistently reflected by the low correlation coefficient r=0.36 for a simple linear correlation. Furthermore, we found that the ultrasound velocities through the aqueous and vitreous do not significantly depend (95%) on the age of the horse.

Discussion



We have calculated conversion factors for the three types of healthy equine ocular media, which are useful to correct the calculations of axial intraocular distances that are based on the speed of ultrasound in the human eye. The speed of ultrasound in human aqueous, lens and vitreous at 37° C are 1532 m/s, 1641 m/s and 1532 m/s respectively [21]. The corrections for aqueous and vitreous are on a sub-percent magnitude and it is therefore safe to neglect these in a clinical setting. The correction for the equine lens, which is 1.008(18), is on a percent-level and in this respect closer to the human lens than to the values that have been reported for pig, dog and rabbit lenses, which are 1.024, 1.040 and 1.055 respectively [10].

The present study shows that the speed of ultrasound through the optical media of the equine eye does not change during the first 72 hours after enucleation. In a previous study in human eyes it was concluded that there is no measurable change in the velocity of ultrasound in ocular media during the first few days after death [9]. Another study concerning the structure and regional water content of lenses in humans, cattle and pigs concluded that there are no changes in structure and water mobility within the first 3 days after death [22]. Also the storage did not influence the speed of ultrasound. This was proved by our statistical analysis. As well in a previous study, which analyzed the thermal stability characteristics of the lenses of vertebrates, no effects of freezing on stability properties of the lens were observed [23]. We therefore assume that our results closely resemble the ocular ultrasound velocities in the living animal.

For the present study a 4 MHz transducer was used throughout, which is not the commonly used frequency for ocular ultrasound measurements. Axial length measurements with A-mode are generally conducted with a frequency of 8 MHz and B-mode ultrasonography with a frequency of 10 MHz [3]. In previous studies a range of different ultrasound equipment has been used, operating at frequencies of 4 MHz [9], 10 MHz [2,10,19] or 20 MHz [24]. The differences in the ultrasound velocity for separate frequencies have been described to be less than one percent [25]. We are therefore confident that it is sufficient to use a single frequency for all measurements.

Previous studies have shown that the human central corneal surface temperature is 34.8° C, which is 2.2°C lower than the average physiological body temperature [26]. Other studies have indicated that the temperature within the globe increases from the cornea towards the more posterior structures [26]. The anterior lens surface temperature in rabbits has been shown to be approximately 36°C, whereas the posterior lens surface was 38°C [10]. Since the physiological body temperature of the horse varies between 37.5°C and 38.2°C, we assume the equine corneal temperature to be in the range of 35.3°C to 36.0°C.

As visible in Table 1 the velocity of ultrasound in equine aqueous and vitreous matches values for a broad

Page 4 of 6

Page 5 of 6

Study	Species	Temp.	c aqueous (m/s)	c vitreous (m/s)	c lens (m/s)
Jansson and Kock [9]	Man	37℃		1532 ±0.5	1641 ±1.2
Oksala & Lehtinen [13]	Cattle	22°C	1495	1495	1650
Coleman, Lizzi [21]	Man		1532	1532	1641
Görig [10]	Dog	36°⊂		1532 ±0.7	1707 ±19
Görig [10]	Rabbit	36°⊂		1531 ±1.1	1731 ±21
Görig [10]	Pig	36°⊂		1532 ±0.8	1681 ±6.3
Osuobeni and Hamidzada [11]	Camel	20°C			1686 ±16
Hamidzada and Osuobeni [27]	Camel	20°C	1499 ±23	1497 ±24	
Present study	Horse	36°⊂	1529 ±10	1527 ±16	1654 ±29
Coleman, Lizzi [20]	Man Cataract	37°C			1629 ±38

Table 1 Compilation of results from the literature for the ultrasound velocity in different ocular tissues

range of different species like human, bovine, canine, rabbit, porcine and camel [9,10,13,21,27]. Both values lie slightly above the ultrasound velocity of distilled water, which is 1521 m/s at 36°C, and this further supports the hypothesis that the speed of ultrasound in aqueous and vitreous is dominated by its water content for all species. Consequently, we found that the speed of ultrasound in these ocular tissues increases with increasing temperature, which has been described for other species as well [2,10]. In contrast, our results for the velocity of ultrasound in equine lenses do not indicate a temperature dependence. However, another study reports an increase of the speed of ultrasound in porcine ocular lental tissue with rising temperature [12]. The considered temperature range in this study is greater (23°C - 37°C) than in the present study. Yet, the regression formula c =1642,61 +1,00 x T (m/s) shows that the temperature dependence is very small. It is known that the speed of ultrasound of a tissue highly depends on its content of protein and water. The temperature dependence of ultrasound is mainly determined by the respective dependence of the compressibility. For water, the compressibility decreases for increasing temperature up to a minimum at 67°C, i.e. the ultrasound velocity increases. For all other liquids however, the ultrasound velocity decreases and the sign of the temperature coefficient is different. The human lens consists of about 66% of water and 34% of protein [22]. Concerning lental proteins, two groups can be distinguished, soluble and insoluble, and the ratio of these vary with species and age [3]. The water content of the lens, and thus the velocity of ultrasound (Table 1), also varies among species [22] and lenses are therefore expected to show a different behaviour with changes of temperature.

In general, the amount of water in the lens consists of free and bound water. The ultrasound velocity of free water is lower than of bound water. The cortex has the highest percentage of free water, whereas the nucleus consists of a low amount of free water and a high amount of protein-bound water [22]. Hence, it can be expected that the speed of ultrasound in the lental periphery (cortex) is slower than in the nucleus [24]. The equipment used in this study was not suited to resolve these differences and we determined the ultrasound velocities in average for the lens as a whole.

The amount of free water in the human lental nucleus increases during the aging process [22,28] and a decrease of the ultrasound velocity for older horses may therefore be expected. However, in the present study, no significant correlation (95%) between age and velocity of ultrasound for horses between 6 and 22 years of age could be determined. Due to a lack of samples, no conclusion can be drawn for horses older than 22 years. In a study of human lenses with a wider age distribution, a correlation between the velocity of ultrasound and age was found [20].

In a previous study it was reported that the speed of ultrasound is slower in cataractous lenses, compared to healthy lenses, because of the increased water content [20]. Another study describes that during cataract formation the content of bound water in the lental nucleus is decreasing in favour of "free" water [22]. Therefore, the ultrasound velocity decreases as well. For a correct determination of the axial ocular dimension in equine cataractous eyes further evaluations of such eyes are necessary.

Conclusions

The main result of this study is the calculation of the conversion factors which are essential to correct axial ocular dimensions measured with ultrasound equipment calibrated for human use. We found that the conversion factors for the aqueous and vitreous can safely be ignored. However, for the equine lens, the ultrasound velocity is significantly higher (95%) than in the human lens and measurements must be corrected.